TECHNICAL DATA SHEET 500-1000-407

Date of issue: 09.06.2015

Product Code: 500-1000-407

Description



5ml Polystyrene Bijou, Plain, Gamma Irradiated. Product supplied in 1000's bagged in 1000's

Declaration of Conformity

This product comes under the remit of the In Vitro Medical Devices Directive 98/79/EC and is classed as a general IVD product under Annex III (specimen receptacle as per article 1(b) of the directive).

As the manufacturer we hereby declare that the requirements and obligations, as described in Annex III sections 2 to 5 of the Directive have been met.

Signed on behalf of Elkay Laboratory Products Ltd

R. Conway

Managing Director

Authorised by: R. Conway

Elkay Laboratory Products (UK) Ltd.
Unit E, Lutyens Industrial Centre, Bilton Road, Kingsland Business Park, Basingstoke RG24 8LJ, United Kingdom
tel. 01256 811118 fax. 01256 811116 sales@elkay-uk.co.uk www.elkay-uk.co.uk
Company Registration Number: 288 4927 Registered Address: as above VAT Number: GB614 8952 20

Microbiological Status	Gamma Irradiated 25-40 Kgy
Standard to which the product conforms	BS EN 14820:2004
Material of construction (body)	Polystyrene
Material of construction (Cap shell)	Polypropylene
Volume (ml)	5ml to upper graduation, 7ml to brim.
Bottle Dimensions (mm)	55.05 x 17.25mm (Body) (approx)
Operational temperature Range	0 to 65 Degrees Centigrade
Inner Pack Quantity	1000
Outer Box Quantity	1000

Authorised by: R. Conway

Elkay Laboratory Products (UK) Ltd.
Unit E, Lutyens Industrial Centre, Bilton Road, Kingsland Business Park, Basingstoke RG24 8LJ, United Kingdom
tel. 01256 811118 fax. 01256 811116 sales@elkay-uk.co.uk www.elkay-uk.co.uk
Company Registration Number: 288 4927 Registered Address: as above VAT Number: GB614 8952 20